

BOSE INSTITUTE KOLKATA		
Tender No.	:	R/PCS/DMM/MP/Web/7
Tender date	:	18.11.13
Type of Tender	:	Limited
Tender Title	:	<u>UCD-200 TO Bioruptor Standard Sonication device with 1.5 & 15 ml tube holder & new soundproof box.</u>
Specification	:	Application suitable for chromatin shearing, DNA shearing, cell lysis Ultrasonic wave frequency~20KHz Power Supply: 230V, 50/60 Hz, Sonication Unit dimensions: 175(W)X160(D)X265(H) mm (Max) Sound proof box dimensions: 350(W)X350(D)X500(H) mm (Max) Water bath volume ~ 750 ml Timer, Analog Max 15 min. Total weight~30 Kg Tube holder of capacity 0.5-1.5-10-15 and 50 ml Minimum number of samples to be processed simultaneously 0.5 ml tubes-12; 15 ml tubes-6; 10 ml tubes-6; 1.5 ml tubes-6; 50 ml tubes-3.
Quantity	:	One unit.
Last date & time for submission	:	10.12.2013 upto 2-00 p.m.
Date & time for opening of bids	:	10.12.2013 at 3-00 p.m.
Submission of Tender (Address)	:	Division of Molecular Medicine Bose Institute, P-1/12, C.I.T. Scheme VII M, Kolkata 700054
Venue of bid opening	:	Office of the Division of Molecular Medicine
For any query the interested bidders may contact (Dept./Section/Div./Unit)	:	Office of Division of Molecular Medicine Bose Institute, Kolkata 700054 Ph: 033 2569 3256

General Terms & Conditions ::		
Warranty	:	Should be mentioned
Payment terms	:	Payment will be made after complete delivery of the instrument in good condition and satisfactorily installation
Delivery schedule	:	Within 8 weeks from date of order & if any defect of the supplied item is found, it should be replaced immediately from your side.
Bid security (earnest money deposit), if applicable	:	2% of the quoted value
Submission of Performance Bank Guarantee (PBG), if applicable	:	10% of the order value, for the successful bidder.
Any other information (if applicable)	:	-
<i><u>Name of the instrument and submission of tender should be mentioned on the envelope positively.</u></i>		
Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.		

Registrar, B.I.