BOSE INSTITUTE					
KOLKATA					
Tender No.	:	R/PCS/DMM/KB/Web/1			
Tender date	:	DATE: 27.05.2014			
Type of Tender	••	Limited			
Tender Title	••	Gradient PCR machine			
Specification	:	• Display: High resolution color touch screen			
		• Ramp rate: - Block ramp rate not less than 6 C/sec			
		Sample ramp rate in between 4-5 C/sec			
		• Block conformation: 3x32 formats or 2 x 48			
		formats System should be able to accept multiple			
		and different types of blocks			
		• Sample capacity: 2 x 48 x 0.2ml tubes or			
		3x32x0.2ml			
		• Temp. Range: 0-100 C			
		• Memory : >1,000 typical programs onboard ;			
		unlimited with USB flash drive			
		• Initial PCR start up reagents to be supplied with			
		the thermal cycler			
Quantity	:	1 (One)			
Last date & time for submission	:	27.06.14 upto 14.00 hrs.			
Date & time for opening of bids	:	27.06.2014 at 15.00 hrs			
Submission of Tender (Address)	:	Division of Molecular Medicine			
		Bose Institute, P-1/12, C.I.T. Scheme VII M,			
		Kolkata 700054			
Venue of bid opening	:	Office of the Division of Molecular Medicine			
For any query the interested bidders	:	Division of Molecular Medicine			
may contact		Bose Institute, Kolkata 700054			
(Dept./Section/Div./Unit)		Ph: 25693217			
		Ph. 25693321			

General Terms & Conditions ::		
Warranty	:	At least one year from the date of installation
Payment terms	:	After delivery and successful installation along-with submission of bill
Delivery schedule	:	Within 8 weeks from date of order
Bid security (earnest money deposit), if applicable	:	NA
Submission of Performance Bank Guarantee (PBG), if applicable	:	10% of the order value, to be submitted by the successful bidder at the time of award of the contract in the form of A/c payee demand draft/Bank Guarantee from any nationalized bank in acceptable form to remain valid for a period of sixty days beyond the date of

	completion of all contractual		
	obligations of the supplier including		
	warranty obligations.		
<u>Name of the instrument</u> and <u>submission of tender</u> along with tender no and date should be mentioned on the			
envelope positively.			
Director Bose Institute reserves the right to accept or reject any	or all tenders either in part or in full. The reason		

Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.

Registrar