BOSE INSTITUTE KOLKATA

Tender No.	DPB/DST-SERB/DNSG/04/02-2015	
Tender Date	17.02.15	
Tender Type	Limited	

Tender Title	Hybridisation Shaker Incubator with Three bottles. 220 voltage	
Specification	 Compact bench top incubator for multiple temperature-sensitive applications, ranging from incubation of overnight cultures to nucleic acid hybridization. With touch-screen control panel and easy-to programme 3- line display. Should be 32-bit microprocessor control with a speed of 0- 200 rpm. 	
	 Temperature should not be less than Amb+5°C to 85°C. Temperature uniformity should be ±0.2°C and resolution should be 0.1°C. Space-saving stackable design with Wipe-clean painted metal chassis and stainless steel interior. Should have programmable timer 1-9999' with alarm. Should come with bidirectional orbital shaking platform for clockwise and anticlockwise motion. 	
	 Should have Safety-door & thermal safety switches; auto shut-off upon fan failure; auto-recovery after power failure. Options for flask holders and rotisserie should be available. Three Bottles (one large, one small, one-extra small) for hybridization, with one suitable Voltage stabilizer with Back up for 1-2 hrs. 	
Quantity	One with 3 bottles, connector, and a voltage stabilzer	

Last Date & Time for submission	18.03.2015 upto 14.00 hrs
Date & Time for opening bids	18.03.2015, at 15.00 hrs
Submission of Tender (address)	Prof in Charge, Division of Plant Biology, Bose
	Institute,93/1 A.P.C.Road, Kolkata 700 009
Venue of bid opening	Same as above
For any query the interested bidders may	Prof. Dibyendu N. Sengupta, Division of Plant
contact (Dept./Section/Div./Unit)	Biology, Main Campus, 033-23031130

General Terms & Conditions

Warranty	1 year with back up service from the company
Payment terms	After successful delivery and installation of the
	instrument
Delivery schedule	Within 30 days of receipt of purchase order
Bid security (earnest money deposit) if	NA
applicable	
Submission of Performance Bank Guarantee	NA
(PBG), if applicable	
Any other information (if applicable)	NA

Name of the instrument and tender no. should be mentioned on the envelope positively

Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.

Sr. Professor & Incharge, Registrar's Office